SHORT REPORTS

Failed laparoscopic clip sterilisation

Laparoscopic sterilisation using Hulka-Clemens spring-loaded clips has been used extensively in Britain since the method was described by Hulka *et al*¹ and Lieberman *et al*.² We here describe a case in which this technique failed despite correct application of the clips.

Case report

The patient, aged 35, para 3+0, had a vaginal termination of pregnancy at seven weeks and laparoscopic clip sterilisation, one clip being applied to each uterine tube 1 cm from the cornu. Care was taken to ensure that each clip was completely across the tube with the spring rammed home. No part of the round ligament, ovarian ligament, or mesosalpinx was included in the jaws of either clip. Fifteen weeks later the patient was found to be nine weeks pregnant. She had a further vaginal termination of pregnancy and bilateral partial salpingectomy through an abdominal incision. Each clip was completely across its tube but histological examination of each tube showed an oval lumen at the central area of clamping of a minimum diameter of 0.0125 cm on the left side and 0.00125 cm on the right side. There was epithelium lining each lumen but no plicae (figure). Several sections were taken from each tube and in each the lumen was continuous with the lumen of the uncrushed parts, so that no part of either tube had been completely occluded.

Photomicrograph of transverse section through mid-point of crushed segment of each tube. $\times 315$ (original magnification).

Comment

A fertilised human ovum at the four-cell stage, the stage of development at which it passes through the isthmic portion of the tube, is 0.013 cm in diameter.³ It might therefore be able to pass through a

lumen of 0.0125 cm, especially as there was probably some shrinkage of the lumen during preparation of the specimen. Each lumen was clearly wide enough to allow the passage of spermatozoa. We therefore wonder whether patients who have had a laparoscopic clip sterilisation may be prone to ectopic pregnancies. Such a case has been recently described.⁴ Laparoscopic clip sterilisation is clearly safer than laparoscopic tubal electrocoagulation, but to minimise the overall failure rate the clips should probably not be applied at the same time as terminating a pregnancy. The uterine tubes are more difficult to find when the uterus is bulky and more oedematous than in the non-pregnant state. It is more difficult to place the clip completely across the tube, and more difficult to ensure that it is applied at a right angle to the long axis. Apparently even when the clips are correctly applied the uterine tubes are not always completely occluded.

¹ Hulka, J F, *et al*, *American Journal of Obstetrics and Gynecology*, 1973, **116**, 715.

² Lieberman, B A, Bostock, J F, and Anderson, M C, *British Journal of Obstetrics and Gynaecology*, 1974, **81**, 921.

³ Jeffcoate, T N A, *Principles of Gynaecology*, 4th edn, p 101. London, Butterworth, 1975.

⁴ Clarke, G A, Letchworth, A T, and Anderson, M C, *British Medical Journal*, 1979, **1**, 659.

(Accepted 6 July 1979)

Westminster Hospital, London SW1P 2AP

A KENNEY, FRCS, MRCOG, senior gynaecological registrar

R D DE VERE, FRCS, FRCOG, consultant gynaecologist

J R JASS, BSC, MB, registrar in histopathology

Effects of the media on attitudes to electric convulsion therapy

Some doctors are afraid that the increase in media coverage of electric convulsion therapy (ECT)¹ might adversely affect patients' attitudes to what many consider to be an effective treatment.² In this study I assessed these effects on patients treated with ECT soon after maximum media coverage; I also gained other information about patients' knowledge and fears of ECT.

Patients, methods, and results

I interviewed all patients who had ECT between December 1977 and April 1978 at one hospital (Barrow Hospital, Bristol). Three patients were unable to complete the interview. Thirty patients remained (11 men, 19 women, age range 31-86); 28 were depressed and two schizophrenic. Eighteen had had ECT before. A partially structured interview was held with each patient on the day before they were to start ECT but after they had given consent. To avoid any change in the normal procedure, medical staff were not informed of the interview.

Most patients (26/30) felt satisfied about why they were having ECT, although four did not. The hospital doctor was the main source of information for 21 patients and mentioned by 25. Ten patients mentioned other patients and one mentioned nurses as a source of information. All patients knew that they would have an anaesthetic; 18 knew that they would have an electric shock, and nine that they would have a seizure. Fourteen had some idea of the duration of the anaesthetic (3-10 minutes): estimates ranged from four seconds to four hours. The aspect most commonly feared was the anaesthetic: 15 mentioned this and 12 feared this most. Personality change was feared by 10, memory loss by six, and dying by five. Six did not feel fearful.

Eleven patients were aware of publicity about ECT. Only one person felt that her attitude towards ECT had been affected by this; she was also the only person who felt that the publicity was exaggerated. The table compares those aware of publicity with those not aware. The two groups did not differ in age, sex, social class, education, or experience of ECT. There were no significant differences between those who had had experience and those who had not in knowledge of or fears about the procedure.