ABC of ENT

HAROLD LUDMAN

HELP FOR DEAF ADULTS

If more than slight, deafness holds serious implications for the patient's family, social, and working life. Difficulties with a job may be exacerbated by the need to use a telephone. These generate financial worries, which, together with the problems of sharing activities such as watching television or going to the theatre, create tensions in the home. The deaf patient becomes isolated and threatened and, not surprisingly, depressed. A major source of help is to provide a hearing aid.

Getting a hearing aid

As a rough guide anyone whose pure-tone audiogram in the better ear shows an average loss of 40 dB or more in speech frequencies (500, 1000, and 2000 Hz) is likely to have sufficient conversational difficulty to justify trial of a hearing aid.

Various hearing aids, both body worn and behind the ear, covering most needs are available through the NHS. The GP should refer the patient to the ENT department of a hospital with facilities for distributing hearing aids. He will be assessed in the outpatient department before being fitted with an aid.

Alternatively any patient may seek advice directly from a hearing aid dispenser. Their controlling body—the Hearing Aid Council—keeps a register of dispensers. If a hearing aid dispenser discovers that a patient might have conductive deafness or features other than those of a slowly progressive bilateral sensorineural loss he has by law to refer the patient to his GP. Many commercial dispensers will let a patient have an aid on trial for a time before purchase, and they have to make clear the terms of the trial. Patients should take advantage of such trials.

Types of hearing aid

All electronic aids consist of a microphone, an amplifier, and an earphone. They are powered by dry batteries and have on-off and loudness controls. Some aids offer an induction coil or a T setting, which enables the aid to use electromagnetic induction fields and provides better quality sound from electromagnetic sources such as telephones, televisions, and microphones.

With some aids the amplifier and microphone are worn in a box on the chest with a wire attached to the earphone; in others all the components are in a small package on or near it. The latter aids, which are the more common, usually lie behind the ear with the microphone connected to the earphone by a transparent hollow tube. Some commercial aids incorporate the components in spectacle frames or in a "all-in-the-ear" package.

Aids differ electronically in terms of the gain, maximum output, and frequency response, and in measures to reduce the effects of sudden loud noise. They also differ in the way that sound is conveyed. This is usually to an insert which fits tightly into the meatus on the same side as the aid. This provides air-conducted sound. Alternatively sound may be conveyed to a bone-conducting transducer pressing on the skull behind the ear. Sound may be taken from one side of the head to the other by a CROS (contralateral routing of signals) fitting. Two aids may be fitted, and combinations of transmission from one side to the other provide varieties of CROS fitting.

Choice of aid

For an aid to be useful for speech discrimination the patient needs to distinguish at least half of speech content. Accurate prescription in the way that refractive errors are assessed is impossible. Even so, the unaided speech audiogram—and the presence or absence of recruitment (the phenomenon found with cochlear defects in which loud sounds are abnormally loud and often intolerable) will indicate the characteristics on which to base a choice for “trial and error.”

An ear-level, behind-the-ear aid, with a firmly fitting insert of hard acrylic is the basic form of aid, with various electronic features to suit most patients. A patient who is sensitive to loud noise will need some form of “automatic gain control” to reduce the effect of sudden loud noise. If very high amplification is needed there may be problems with acoustic feedback, producing the familiar high-pitched whistling sound. This occurs whenever the outlet to the ear is not acoustically isolated from the microphone and will affect any aid with an insert that does not fit the ear tightly enough.

High gain that causes feedback despite a properly fitting insert demands that the aid be distant from the earpiece. A body-worn aid may be advisable. Alternatively the aid can be connected by a CROS fitting. Body worn aids, apart from being conspicuous, cause unwanted noise from rubbing against clothes; ear-level aids may be affected by wind noise.

Some moulds cause allergic reactions in the external ear canal and alternative materials must be used. Any condition with discharge from the ear—a mastoid cavity, chronic middle ear disease, otitis externa—will be adversely affected by an airtight plug. One solution is to fit a bone-conducting aid, but this needs greater power and causes discomfort because of the pressure needed to hold it to the head (by a springy headband or special spectacle frame). The other is an open mould, in which just the tip of a tube fits into the ear but does not occlude it, or a CROS fitting, with the aid on the other side to prevent feedback.

Arthritic patients may find the tiny controls of ear-level aids hard to manipulate. Spectacle aids have the disadvantage that repair or loss of either spectacles or hearing aid deprives the user of both.

Which ear to fit?

If the better ear is useful for hearing speech the poorer is usually fitted with an aid. If neither ear gives useful hearing the better ear is generally fitted. Some patients benefit from an aid in both ears. Patients with unilateral loss may occasionally be helped by a CROS fitting from the deaf side to a tube, fitting, but not obstructing the exterior of the canal on the normal side.

Even if an aid does not discriminate speech usefully it may help lip reading, indicating rhythms and cadences of speech; in the worst case it may be useful as a warning of impending danger.

It is a sad irony that hearing aids give best results for patients with conductive deafness, for whom there are often operative alternatives.

A patient may be disappointed once his aid is fitted if he expected nearly normal hearing. No aid provides normal sound, and patients need counselling and auditory training if their aids are not to end life in a drawer. They need to be encouraged and taught to interpret and use the sounds, distorted compared with normal hearing, that the aid can provide.

Old-fashioned speaking tubes or ear trumpets may be surprisingly helpful for the very deaf.

Lip reading

All normally hearing people watch faces as a supplement to hearing, and instruction in lip-reading is essential for any severely deaf patient. If a patient has a progressive hearing loss that will eventually become severe or total, help with lip reading should be sought while the level of hearing is still useful. The availability of instruction in lip reading varies. Advice should be sought from local adult education bodies, or, as with any problem raised by this article, from the Royal National Institute for the Deaf, 106 Gower Street, London WC1. In London the City Lit Centre for the Deaf provides advice and training in all aspects of auditory rehabilitation, lip reading, manual communication, and speech and machine shorthand (palantype) operation. The City Lit also runs a course for hearing therapists.

The effect of sudden severe or total deafness on a previously hearing adult is as apocalyptic as a major stroke. The patient and his family need help from psychiatrists, social workers, and teachers of the deaf. Ideally the patient should have a period of residential care. He must never be left to fend for himself in a silent world.

Link, the British Centre for Deafened People, at Eastbourne, is a registered charity providing residential rehabilitation courses, primarily for adults of working age and the victims of sudden deafness.

Environmental aids

There are many fairly cheap aids that can make life easier for the deaf.

Telephones—Advice and devices are available from telephone sales offices: (a) bells may be replaced with extension bells, buzzers, hooters, or flashing lamps; (b) amplifying handsets provide an adjustable volume control, (c) inductive couplers provide an electromagnetic field which can excite an induction coil in a hearing aid in the T (telephone) position, (d) separate receivers allow a normal-hearing person to hear the call and “lip speak” to a deaf listener.

Watching television is socially less divisive if the deaf patient can adjust his volume independently of the loudspeaker volume of the set. Most television retail or rental firms supply various devices. Television adapters have an isolating transformer that is attached to the set and linked by a wire to a gain control box to which the patient's earphone is connected. Induction loops are wires placed around a room from isolating transformers which generate electromagnetic fields that can be picked up by the induction coil of a hearing aid set in the T position. This provides better quality sound without interference from the amplification of unwanted environmental noise. Microphone aids, attached to the patient's ear through an amplifier, are stuck directly to the grill of the TV set.

Tinnitus

Management of tinnitus

Teletext decoders allow the viewer to receive written information on his screen from Ceefax (BBC) and oracle (ITV). The user can select the information he wants. Oracle offers information on deaf societies as well as instruction in lip reading. A "mix" facility allows text to be shown over the programme on the set, which will be particularly useful when subtitles are more generally available. Prestel, which funnels information from telephone lines to the television set, offers much more information but not specifically for the deaf.

Alarms—Door bells can be modified to buzzers, lights, hooters or gongs. Baby alarms and clock alarms can also activate vibrators or lights.

For many severely deaf patients tinnitus is almost as worrying as the deafness. For some with moderate hearing loss tinnitus may be the dominating problem, while in a few a hearing defect is found only when tinnitus is investigated.

Very rarely tinnitus may be objective—that is, created by sound waves generated within the body by vascular tumours, abnormal blood flow, palatal myoclonus, or an insect in the meatus. The noise is then audible to an examiner. In conductive deafness tinnitus is often the effect of removing ambient masking noise so that body activities become audible.

Tinnitus in sensorineural deafness can be explained by a change in resting activity in the auditory nerve or its central pathways. Everyone experiences tinnitus at times and almost all will hear "white noise" when placed in sound-proofed surroundings. The resting magnitude of spontaneous activity in the auditory system is only just below that at which a sound enters consciousness: this is the price paid for the sensitivity of normal hearing. It is not surprising therefore that many mechanisms in sensorineural deafness may—for example, by affecting inhibitory influences on the cochlea—change the magnitude of resting activity and produce tinnitus. Perhaps it is also not surprising that there is yet no proved drug remedy for this symptom of protean inner ear malfunction. Carbamazepine is often worth trying, and the effects of local anaesthetics are under investigation.

Reassurance—Many patients fear that tinnitus is a symptom of serious brain disease or impending stroke. Careful explanation of the mechanisms is time well spent.

Psychotherapy—There is a relation between tinnitus and depression, whereby the patient may blame the depression on the tinnitus, when in fact the relation is complex and two way. If depression can be identified and relieved tinnitus becomes much less troublesome.

Masking sound—Many patients find ambient noise helpful. At night this can be provided by the white noise of a radio tuned "off station." Tinnitus maskers, which look like hearing aids and produce different types of noise, are sometimes useful. They can be obtained on the NHS (subject to the resources of the ENT department) and are available commercially. Their value can be assessed only by trial and error; for a few it is considerable.

Cochlear nerve section should be considered only when an ear is totally deaf, as this measure produces relief in only about half the patients. Since tinnitus can be worse afterwards the possibilities and risks must be discussed carefully with a psychologically stable patient. This procedure has only a limited scope.

The photographs of the spectacle and "within-the-ear" aids were reproduced by permission of P C Werth Ltd and that of an amplifying telephone handset by permission of the Post Office.

Mr Harold Ludman, MA, FRCS, is consultant laryngologist, King's College Hospital, and neuro-otological surgeon, the National Hospital for Nervous Diseases, London. This is the sixth of 16 articles.