basic district RAWP target, giving each health authority a clear picture of the resources available for its own population. Cross boundary flows might then be agreed among districts through discussions that should fully include local general practitioners, and authorities might seek to reduce or maintain the flows according to their own preferences. Secondly, family practitioner committee resources should be included within the total district allocations—at least for planning purposes if not for administration—to clarify the relation between total NHS expenditure and levels of health. Regional health authorities would then, at last, be able to measure and compare the efficiency and effectiveness of districts, supplementing the limited measures that are available at present and moving towards real equality of health care.¹⁰

The government has recently requested the NHS Management Board to review the operation of the RAWP formula, taking into account the patterns of illness across the country and the special problems of the inner cities.¹¹ This

review provides an opportunity to achieve this aim of equality.

MARK MCCARTHY

Senior Lecturer in Community Medicine,
School of Medicine,
University College London,
London WC1E 6JJ

- 1 Department of Health and Social Security. *Sharing resources for health in England. Report of the Resource Allocation Working Party*. London: HMSO, 1976.
- 2 Dearden RW. *Resources and health deprivation*. Birmingham: University of Birmingham Health Services Management Centre, 1985. (Discussion paper No 19.)
- 3 Department of Health and Social Security. *Report of the advisory group on resource allocation*. London: DHSS, 1980. (Mimeograph.)
- 4 Paton C. *The policy of resource allocation and its ramifications*. London: Nuffield Provincial Hospitals Trust, 1985. (Occasional papers 2.)
- 5 Bevan G, Brazier J. Subregional RAWP—Hobson's choice? *Hospital and Social Service Journal* 1985;94:1064-5.
- 6 Bevan G, Beech R, Craig M. Alternatives to RAWP. *Health and Social Services Journal* 1985;94:1098-9.
- 7 Bevan RG. Organising the finance of hospitals by simulated markets. *Fiscal Studies* 1984;5:44-63.
- 8 Anonymous. Applying economics. *Health Economics Bulletin* 1985;5:3-5.
- 9 McCarthy M. A five-year plan for better health. *The Health Services* 1982 June 11:14-5.
- 10 McCarthy M. Epidemiology—a losing cause? *Br Med J* 1983;286:1682-3.
- 11 Anonymous. Resource distribution 1986-7: RAWP formula to be reviewed. *Br Med J* 1986; 292:74.

Cobalt on the way out

Two types of megavoltage apparatus are found in most British radiotherapy departments: telecobalt units and linear accelerators. Although the Department of Health and Social Security has stated that linear accelerators may be considered to be suitable cobalt substitutes for departments of all sizes,¹ the demand for cobalt units was widely expected to continue. In practice, few departments now seem prepared to consider a new telecobalt unit, and the move to linear accelerators is general. Why should this be?

The term megavoltage as used in radiotherapy means beams of x or γ ray photons with energies above one million electron volts (1 MeV); the same term may be applied to the equipment generating beams of these energies. Megavoltage beams are more penetrating than those at lower energies and so are useful for treating tumours at depth below the skin. In addition, these beams tend to reduce the skin dose, thus minimising one complication of therapeutic irradiation.

Radioactive cobalt is an isotope produced in a reactor and has a half life of 5.2 years; it decays to form nickel with the emission of photons (γ rays) of 1.17 and 1.33 MeV. Used as a teletherapy source cobalt produces a beam quality comparable with that of a 2-3 MeV linear accelerator. Cobalt units offer a simple approach to the production of a megavoltage beam and require a minimum of servicing and technical support. The output of the machines is low and decreases as the cobalt decays. This would make treatment times unacceptably long, so that in practice the source is replaced every three to four years at a cost of £25 000-£30 000—an expense that needs to be included in the overall costs of running such a unit.

Modern linear accelerators operating at 4-6 MeV are much more complicated and need a great deal more technical support than cobalt units. The beams generated by these machines are of a higher mean energy than those from cobalt and are more penetrating. Large areas can be treated without the need for multiple fields. Output rates are three to four

times greater than those of cobalt units permitting up to twice the number of patients to be treated daily. So, while linear accelerators cost more initially, estimates of unit treatment costs show a favourable comparison with cobalt units.²

Furthermore, the use of cobalt beams for the treatment of more deeply seated tumours may give less good results than those obtainable from linear accelerators. The difference may be linked with the observation that units equipped solely with cobalt tend to lack the supporting facilities necessary for the planning and delivery of the highest quality of service. If there is a discrepancy between the results obtainable the explanation is multifactorial. To suggest that any difference is due to a difference in quality between cobalt and accelerator beams is an oversimplification.³

Very few cobalt units are now available commercially, a lack of choice that reflects a rapidly shrinking market. Most National Health Service radiotherapy units have two megavoltage units and cobalt might appear to offer a reasonable "second" machine, but the higher working rate (and so the greater patient capacity) of linear accelerators dictates a choice in their favour. The past decade has seen an increase in the workload of most radiotherapy departments out of proportion to the changes expected as a result of an aging population. This increased load can be met only by machines capable of a higher workload than cobalt teletherapy units; they have, like steam engines, been made obsolete by technical advances.

ROGER BUCHANAN

Consultant in Radiotherapy and Oncology,
Wessex Regional Radiotherapy Centre,
Royal South Hants Hospital,
Southampton SO9 4PE

- 1 Department of Health and Social Security. *Health circular HC(78)32*. London: HMSO, 1978.
- 2 Greene D. The cost of radiotherapy treatments on a linear accelerator. *Br J Radiol* 1983;56:189-91.
- 3 Hanks GE, Diamond JJ, Kramer S. The need for complex technology in radiation oncology. *Cancer* 1985;55(suppl):2198-201.