SHORT REPORTS

Measurement of urea as a side ward investigation

Chemical pathology originated in the side ward, but increased sophistication of investigations and a higher demand for them led to the establishment of central, purpose built laboratories. The design of some analytical equipment has since been simplified, and it is now feasible for investigations to be performed once again in the side ward. A recent criticism has been that a bedside index of renal function has not been available. We assessed the acceptability to users of a reflectance meter (Reflomat, Boehringer Mannheim, West Germany) that measures plasma urea concentration.

Materials and methods

The reflectance meter method is simple and rapid to use. Measurement of urea concentration is based on the use of solid phase reagent strips. Plasma (10 µl) is placed on the reaction layer, which is impregnated with urease. Ammonia from the enzymatic breakdown of urea diffuses through a gas permeable mesh into an indicator layer, causing a blue discoloration. The intensity of the colour (proportional to the urea concentration) is measured after 10 minutes' incubation at room temperature in the reflectance meter.

The method was first evaluated in the laboratory by skilled staff, and performance was satisfactory. Plasma urea concentrations could be measured accurately and precisely in the range 3-20 mmol/l (18-121 mg/100 ml). The coefficient of variation between batches was acceptable (3.4% at 11 mmol/l (66 mg/100 ml)), and results with the meter showed good correlation with those obtained with laboratory equipment (Technicon SMAC II) ($r=0.992$, $p<0.001$).

Comparison of urea concentration measured with Reflomat (doctors) and with SMAC autoanalyser (routine laboratory).

Conversion: SI to traditional units—Urea: 1 mmol/l $\approx$ 6 mg/100 ml.

Junior hospital doctors from one acute medical team were trained to use the method. They then used it over eight weeks to analyse plasma samples on the ward; duplicate samples were saved for non-urgent analysis in the main laboratory. Analysis was performed on 66 samples, and although the doctors were not as precise as the laboratory staff, the coefficient of variation between batches was satisfactory (11.6% at 11 mmol/l (66.1 mg/100ml)). Comparison of the doctors' results with those from the main laboratory gave a correlation coefficient of 0.91 (figure) and showed that under field conditions the reflectance meter could satisfactorily discriminate between normal and impaired renal function.

Comment

The role of laboratories performing emergency tests that are run by doctors is contentious. Many clinicians, especially those who make heavy demands on the chemical pathology service, see advantages in having results more quickly and frequently than a central laboratory can provide. The method described uses solid phase reagent strips and a reflectance meter. It

is similar to techniques that clinicians are familiar with—namely, those for measuring blood glucose concentrations—but requires more input as it entails centrifugation, pipetting, and timed incubation. Despite this a small group of junior hospital doctors admitting patients to general medical wards found it easy to use and were able to screen satisfactorily for renal failure requiring immediate attention.

The system was cost effective. The instrument cost £250, and the cost of the reagent for each test was 28p. If allowance is made for staffing costs then the overall cost of each test of about £1.40 compares favourably with an on call cost for measurement of urea concentration by a central laboratory of £8.50.¹ Recently Boehringer Mannheim produced a new instrument that supersedes the Reflomat, in that it is self calibrating, uses whole blood, and times the incubation automatically. If its efficiency is proved in clinical use this new instrument will make the estimation of blood urea concentration on the wards almost as simple as the estimation of blood glucose concentration.

We are grateful to Dr M Peacock and the junior staff of Leeds General Infirmary for their help and cooperation.

1 Morgan DB, Gunn IR, Faye S, Clegg G, Grant AM. A comparison of alternative arrangements for an out of hours chemical pathology service. *Lancet* 1985;i:859-62.

(Accepted 12 February 1986)

Department of Chemical Pathology, Old Medical School, University of Leeds, Leeds LS2 9JT

S A FAYE, BSC, MSC, senior biochemist

D B MORGAN, MB, MRCPATH, professor of chemical pathology (Professor Morgan died on 18 April 1985)

Department of Medicine, St James's Hospital, Leeds LS9 7TF

P BANNISTER, MB, MRCP, research fellow

Correspondence to: Ms Faye.

Vivax malaria acquired in Trinidad, a malaria free area

I describe a case of malaria caused by *Plasmodium vivax* acquired in Trinidad, which has been declared free of malaria; this is the first case to be reported from there since eradication was achieved. The patient presented in the United Kingdom with an imported fever.

Case report

A 33 year old white man presented after five weeks of intermittent fever and rigors. He had been unable, because of malaise, to complete his daily marathon training of running several miles. He farmed a densely forested area of Trinidad, where he had lived since infancy. Mosquitoes were common in the area and included vectors that transmit malaria (*Anopheles bellator* and *A. aquasalis*). Travel outside the Caribbean had been limited to a visit to Florida two years previously; within the Caribbean he had visited St Martin 18 months previously and Barbados seven months previously, but he had not been to Haiti or the coast of South America. A week before presentation he had flown by standard flight from Trinidad to London with no stopovers. Before departure he had been ill for four weeks.

On admission he was drawn, febrile (temperature 39°C), and suffering from rigors. His spleen was enlarged (2 cm below the costal margin). Blood films showed large solid trophozoites and preschizonts consistent with vivax malaria. Chloroquine 1500 mg was given over two days with primaquine 15 mg daily for two weeks, resulting in a cure.

Intensive epidemiological work was performed, including interviews with contacts and examination of blood films from contacts and people who lived near the patient who were or had been febrile in recent weeks. Blood films proved negative. No malaria antibodies were detected in these subjects. The source of the infection remained puzzling.

Comment

As the result of a programme of eradication from 1957 to 1972 the World Health Organisation in 1982 declared the Caribbean, apart from Haiti and the Dominican Republic, free of malaria.¹ Malaria had been eradicated from

Trinidad in 1965, though the malarial vector, the anopheles mosquito, has apparently remained or returned. I did not think, therefore, that malaria could be a valid differential diagnosis.

Malaria acquired at an airport² was unlikely as the patient had been ill for four weeks before departure and had not left his home in central Trinidad during that time; there had been no stopovers during his flight to London. His visit seven months previously to other Caribbean islands free of malaria was unlikely to have resulted in his acquiring vivax infection, though a six or seven month hypnozoite (exoerythrocytic prepatent) stage could not be ruled out completely. He did not have a history of blood transfusion.

One possible explanation for his infection is that *P. vivax* had been imported from South America. Trinidad is only seven miles from the coast of South America, and South American fishing boats sometimes dock in Port of Spain. Infected anopheles mosquitoes could have been conveyed the short distance over the sea; this hypothesis is similar to that put forward to explain acquired malaria in Europe.² Alternatively, a visiting boatman may have had malaria, and subsequent blood meals taken by uninfected but "receptive" mosquitoes indigenous to Trinidad would have provided the means of importation of vivax malaria.

This case indicates that in an area receptive for malaria where eradication has been achieved the disease may still recur. It further indicates that malaria should be among the differential diagnoses of fever imported to the United Kingdom by travellers from countries free of the disease that are close to, but topographically isolated from, countries in which malaria is endemic.

I am grateful to Dr George Wyatt, senior lecturer in tropical medicine at the Liverpool School of Tropical Medicine, for confirming the presence of *P. vivax* in the blood films, and to Dr Rosemarie Paul, principal medical officer in epidemiology at the Ministry of Health and Environment, Trinidad, for intensive epidemiological work there.

1 World Health Organisation. Malaria risk in international travel. *Weekly Epidemiological Record* 1984;59:221-8.

2 Holvoet G, Michielsens P, Vandepitte J. Airport malaria in Belgium. *Lancet* 1982;ii:881-2.

(Accepted 30 January 1986)

Regional Department of Infectious Diseases and Tropical Medicine, Monsall Hospital, Manchester M10 8WR

M E ELLIS, MRCP, DCH, consultant physician

Absence of seroconversion for HTLV-III in haemophiliacs intensively treated with heat treated factor VIII concentrate

Haemophiliacs are at increased risk of the acquired immune deficiency syndrome (AIDS). The probable causative factor, human T lymphotropic virus type III (HTLV-III), is believed to be transmitted via the clotting factor preparations used by these patients; in particular, commercial factor VIII concentrates made from large plasma pools and obtained from the United States carry a high risk of transmission of HTLV-III.¹ Because retroviruses seem to be sensitive to heat inactivation² heat treated factor VIII concentrates are supposed to reduce the risk of transmission.³

Patients, methods, and results

We studied the rate of seroconversion for HTLV-III in 18 patients with severe haemophilia A (mean age 28, range 15-43). This group comprised all our haemophiliacs who had been given exclusively commercial factor VIII concentrates for home treatment since 1978. Before June 1983 they had received non-heat treated concentrates from the USA (Hemofil, Hyland; Factorate, Armour) for 10-63 months (mean 46 months) at a mean monthly dose of 4073 IU (range 1508-9329 IU). From June 1983 all patients were treated exclusively with a heat treated concentrate from the USA (Hemofil T, Hyland) at a mean monthly dose of 4202 IU (range 469-9030 IU).

Blood was sampled four months before and just before the first infusion of heat treated concentrate and four, 10, 22, and 28 months afterwards. All samples were tested for HTLV-III antibodies with an enzyme linked immunosorbent assay (Vironostika, Organon). Samples that were positive on assay, as well as the first serum sample from each patient, were subjected to immunoblot analysis. The assay and immunoblot analysis were performed by the Academic Medical Centre (Amsterdam, The Netherlands). Immunoblot analysis was also performed by the Centers for Disease Control (Atlanta, USA) on all samples collected from June 1983 to April 1985. No discrepancies were found between the results at the two centres. The table summarises clinical details and serological results.

Before the introduction of heat treated concentrate HTLV-III antibodies were found in only one patient. Afterwards none of the patients seroconverted. All

Characteristics of haemophiliacs treated with heat treated factor VIII concentrate (from June 1983), and rate of seroconversion for HTLV-III

Case No	Dose of factor VIII (IU/month) 1983-5	Presence of HTLV-III antibodies					
		Feb 1983	June 1983	Sept 1983	April 1984	April 1985	Sept 1985
1	3819	-	-	-	-	-	-
2	9030	-	-	-	-	-	-
3	3492	-	-	-	-	-	-
4	1463	-	-	-	-	-	-
5	6370	-	-	-	-	-	-
6	5748	-	-	-	-	-	-
7	2337	-	-	-	-	-	-
8	1981	-	-	-	-	-	-
9	1892	-	-	-	-	-	-
10	469	-	-	-	-	-	-
11	5217	-	-	-	-	-	-
12	7260	-	+	+	+	+	+
13	4510	-	-	-	-	-	-
14	5765	-	-	-	-	-	-
15	2910	-	-	-	-	-	-
16	5749	-	-	-	-	-	-
17	3076	-	-	-	-	-	-
18	4517	-	-	-	-	-	-

patients remained free of AIDS or the AIDS related complex. They had no known risk factors for AIDS apart from the treatment with factor VIII concentrates.

Comment

The prevalence of HTLV-III antibodies in our patients was remarkably low (5.5%), although they were treated from 1978 to 1983 with high doses of commercial non-heat treated concentrates. Considerably higher prevalences have been reported in comparable groups of haemophiliacs, ranging from 72-94% in the USA to 53-66% in Europe,^{1,4,5} and have been related to use of commercial factor VIII concentrates.¹ This difference cannot be attributed either to selection of patients or to methodological errors in the tests for HTLV-III antibodies. Thus our patients probably received fewer batches of concentrates contaminated with HTLV-III. The batches of concentrates that were used until June 1983 seem to have been manufactured about two years before their administration. The prevalence of HTLV-III antibodies in haemophiliacs treated with commercial concentrates from the USA was still low in 1981, and a progressive increase in incidence occurred thereafter.^{4,5} Thus delayed exposure to contaminated batches of non-heat treated concentrate in our patients is possible.

Although from 1983 to 1985 seroconversion for HTLV-III occurred in 27% of Dutch haemophiliacs who were treated with a local non-heat treated concentrate (C Breederveld *et al*, unpublished findings), none of our patients seroconverted. This absence of seroconversion is probably associated with the introduction of heat treated concentrate, although donor selection may also have contributed. Our results agree with those in previously untreated haemophiliacs.³ The mean total dose of heat treated concentrate administered in our group, however, was about 12 times higher than in that study.

We conclude that the heat treated factor VIII concentrate that our patients received carries a low or even no risk of transmission of HTLV-III. Longer follow up is needed to establish whether this product protects haemophiliacs against AIDS.

1 Melbye M, Froebel KS, Madhok R, *et al*. HTLV-III seropositivity in European haemophiliacs exposed to factor VIII concentrate imported from the USA. *Lancet* 1984;ii:1444-6.

2 McDougal JS, Martin LS, Cort SP, Mozen M, Heldebrandt CM, Evatt BL. Thermal inactivation of the acquired immunodeficiency syndrome virus, human T lymphotropic virus III/lymphadenopathy-associated virus, with special reference to antihemophilic factor. *J Clin Invest* 1985;76:875-7.

3 Rouzioux C, Chamaret S, Montagnier L, *et al*. Absence of antibodies to AIDS virus in haemophiliacs treated with heat-treated factor VIII concentrate. *Lancet* 1985;ii:271-2.

4 Anonymous. Blood transfusion, haemophilia, and AIDS [Editorial]. *Lancet* 1984;ii:1433-5.

5 Ball SE, Hows JM, Worsley AM, *et al*. Seroconversion of human T cell lymphotropic virus III (HTLV-III) in patients with haemophilia: a longitudinal study. *Br Med J* 1985;290:1705-6.

(Accepted 17 February 1986)

Haemophilia Centre, Division of Haematology, Department of Medicine, University Hospital, 9713 EZ Groningen, The Netherlands

J VAN DER MEER, MD, consultant in haematology
S DAENEN, MD, consultant in haematology
G W VAN IMHOFF, MD, consultant in haematology
J T M DE WOLF, MD, consultant in haematology
M R HALIE, MD, PHD, professor of medicine

Correspondence to: Dr van der Meer.