

## Cross sectional study of conventional cervical smear, monolayer cytology, and human papillomavirus DNA testing for cervical cancer screening

Joël Coste, Béatrix Cochand-Priollet, Patricia de Cremoux, Catherine Le Galès, Isabelle Cartier, Vincent Molinié, Sylvain Labbé, Marie-Cécile Vacher-Lavenu, Philippe Vielh, for the French Society of Clinical Cytology Study Group

### Abstract

**Objectives** To compare the sensitivity, specificity, and interobserver reliability of conventional cervical smear tests, monolayer cytology, and human papillomavirus testing for screening for cervical cancer.

**Design** Cross sectional study in which the three techniques were performed simultaneously with a reference standard (colposcopy and histology).

**Setting** Public university and private practices in France, with complete independence from the suppliers.

**Participants** 828 women referred for colposcopy because of previously detected cytological abnormalities and 1757 women attending for routine smears.

**Main outcome measures** Clinical readings and optimised interpretation (two blind readings followed, if necessary, by consensus). Sensitivity, specificity, and weighted  $\kappa$  computed for various thresholds of abnormalities.

**Results** Conventional cervical smear tests were more often satisfactory (91% *v* 87%) according to the Bethesda system, more reliable (weighted  $\kappa$  0.70 *v* 0.57), and had consistently better sensitivity and specificity than monolayer cytology. These findings applied to clinical readings and optimised interpretations, low and high grade lesions, and populations with low and high incidence of abnormalities. Human papillomavirus testing associated with monolayer cytology, whether systematic or for atypical cells of undetermined significance, performed no better than conventional smear tests.

**Conclusions** Monolayer cytology is less reliable and more likely to give false positive and false negative results than conventional cervical smear tests for screening for cervical cancer.

### Introduction

Liquid based “monolayer” cytology, possibly combined with human papillomavirus testing, is replacing conventional smear tests for cervical cancer screening in several countries (including the United States and Switzerland). However, there is substantial controversy

about whether the new and costly technologies perform better than conventional cervical smear tests.<sup>1 2</sup> We previously compared the cost of monolayer cytology (ThinPrep, CYTYC; MA, USA) and human papillomavirus testing (Hybrid Capture II test, Digene; Gaithersburg, MD, USA) with conventional smear tests.<sup>3</sup> Here we assess the sensitivity and specificity of the three methods. We also examined the value of human papillomavirus testing in women with atypical squamous cells/glandular cells of undetermined significance (ASCUS/AGUS) and the interobserver reproducibility of the interpretation of conventional smears and monolayer cytology.

### Methods

Full details of the study protocol have been published previously.<sup>1</sup> To avoid spectrum (case mix) bias<sup>5</sup> we considered two groups of consecutive women who were either referred for colposcopy because abnormalities had been detected on previous smears or were attending for routine smears at a French public university (n=2) and private practices (n=2). All procedures were carried out by skilled gynaecologists and experienced cytopathologists. Each woman underwent a standard conventional smear test. The remaining material was then used to prepare the monolayer slide and for human papillomavirus testing. To avoid work up bias, all women were evaluated by all three methods (conventional cervical smear tests, monolayer cytology, human papillomavirus testing) and by the reference method (colposcopy<sup>6</sup> followed by biopsy if abnormalities were detected). To avoid review and context biases<sup>7</sup> cytopathologists read the slides blind to the clinical context in addition to routine reading, separately and independently for the three methods. In cases of disagreement slides were read again to reach a consensus conclusion, given if necessary by an independent expert (optimised diagnosis). In a random sample of the women (30%) we assessed interobserver reproducibility of cytological diagnosis with readings blind to context.

Smear abnormalities were classified into five ordered categories (negative, ASCUS/AGUS, low

Département de Biostatistique, Hôpital Cochin, Assistance Publique-Hôpitaux de Paris, Faculté de Médecine Cochin-Port Royal, Université Paris V, Paris, France

Joël Coste  
*professor of medical statistics*

Service d'Anatomie et Cytologie Pathologiques, Hôpital Lariboisière, Assistance Publique-Hôpitaux de Paris, Paris, France

Béatrix Cochand-Priollet  
*assistant professor of pathology*

Laboratoire de Physiopathologie, Département de Biologie des Tumeurs, Institut Curie, Paris, France

Patricia de Cremoux  
*assistant professor of pharmacology*

Centre for Health Economics and Administration Research (CREGAS), INSERM U537-CNRS UPRESA 8052, Le Kremlin-Bicêtre, France

Catherine Le Galès  
*senior economist*

Laboratoire Cartier, Paris, France  
Isabelle Cartier  
*pathologist*

continued over

bmj.com 2003;326:733

Service d'Anatomie et Cytologie Pathologiques, Hôpital Foch, Suresnes, France  
Vincent Molinié  
*pathologist*

Centre d'Anatomie Pathologique, Besançon, France  
Sylvain Labbé  
*pathologist*

Service d'Anatomie et Cytologie Pathologiques, Hôpital Cochin  
Marie-Cécile Vacher-Lavenu  
*professor of pathology*

Service de Cytopathologie et Cytométrie clinique, Institut Curie, Paris, France

Philippe Vielh  
*pathologist*

Correspondence to: coste@cochin.univ-paris5.fr

grade (LSIL) or high grade (HSIL) squamous intra-epithelial lesions, invasive cancer) and the reference standard into four ordered categories (normal colposcopy or negative biopsy result, cervical intraepithelial neoplasia (grades I, II, and III), invasive carcinoma<sup>8</sup>). We used optimised histological diagnoses for the reference standard. We carried out human papillomavirus testing using the cell suspension that remained after monolayer preparation<sup>9</sup> using low risk (types 6/11/42/43/44) and high risk (types 16/18/31/33/35/39/45/51/52/56/58/59/68) human papillomavirus probes.

### Analysis

We compared the sensitivity, specificity, and proportions of unsatisfactory (according to the Bethesda system<sup>8</sup>) or limited slides using the two tailed MacNemar  $\chi^2$  test. The interobserver reproducibility of the readings was assessed with weighted  $\kappa$  statistics.<sup>10</sup>

To assess potential "sampling bias" due to the split sample technique (monolayer being sampled after conventional cervical smears) we repeated statistical analyses in the subsample of women in whom the residual material after monolayer cytology was sufficient for human papillomavirus testing. The results were similar to those for the whole group and are not shown.

### Results

Between 1 September 1999 and 30 May 2000, 2585 women underwent investigation (table 1). Results of human papillomavirus testing were available from the

1785 women for whom there was enough residual material. The proportion of satisfactory slides was higher with conventional smear testing (91%) than with monolayer testing (87%), though the reasons for unsatisfactory or limited smears differed (table 2). Compared with conventional smear tests monolayer testing consistently showed more abnormalities (especially ASCUS/AGUS) (tables 3 and 4).

Conventional smear tests consistently had superior or equivalent sensitivity, specificity, and likelihood ratios than monolayer tests for the detection cervical intraepithelial neoplasia grade I or higher (table 5) and lesions  $\geq$  grade II or higher (table 6). The sensitivity of systematic human papillomavirus DNA testing (high risk) was no higher than that of conventional smear testing and its specificity was much lower for both grades. For human papillomavirus testing only for ASCUS/AGUS, the sensitivity of the paired monolayer/human papillomavirus testing method was not significantly superior to cervical smear testing.

Interobserver agreement was good for conventional smears (weighted  $\kappa$  0.69, 95% confidence interval 0.64 to 0.74) but only moderate for monolayers (0.57, 0.52 to 0.63) (table 7). The ASCUS/AGUS diagnosis with monolayer testing was especially unreliable.

### Discussion

Our results support the superiority of conventional cervical smear testing, whether considered clinical readings or optimised interpretations, low or high grade lesions, or populations with a low or a high incidence of abnormalities. Human papillomavirus testing, systematic or for a diagnosis of ASCUS/AGUS testing, carried out with monolayer cytology was no better than conventional cervical smear testing. The greater reliability of the interpretation of conventional smears rather than monolayer smears is consistent with their better diagnostic or screening performance. Our findings disagree with those of most previous studies.

We ensured that we obtained the reference standard of colposcopy/histology for all women in the study, unlike previous studies that compared monolayer testing with conventional smear testing and that considered concordant positive and concordant negative tests as true positives and true negatives with discrepancies resolved by consensus review.<sup>11–20</sup> In these studies the proportion of verified cases varied between 0.1% and 30%, and the work up bias was substantial, artificially inflating sensitivity and mathematically favouring the test with the higher rate of false positives: the monolayer technique (or human papillomavirus testing). Two other studies either did not find any difference between the methods<sup>21</sup> or performed post hoc subgroup analyses.<sup>22</sup>

### Limitations of study

In this study the cervical smear was prepared before the monolayer. However, a sampling bias favouring the conventional smear is unlikely as there were very few monolayer slides with only a few cells and the results were similar in the subgroup of women in whom human papillomavirus testing was still possible.<sup>23</sup> The rates of unsatisfactory and limited slides were low, which may be due to our selection of skilled physicians. The cytopathologists were also selected according to their interest in

**Table 1** Characteristics of studied samples by population. Values are numbers (percentage) unless otherwise stated

	Women referred for colposcopy (n=828)	Women attending for routine smear (n=1757)
Mean (SD) age (years)	37.8 (11.6)	33.3 (11.1)
French	605 (73)	1579 (90)
Educational:		
No schooling or primary only	74 (9)	66 (4)
Secondary	472 (57)	927 (53)
Higher	276 (34)	757 (43)
Current smokers	337 (41)	542 (31)
Mean (SD) age at first intercourse (years)	18.5 (2.7)	18.4 (2.1)
Previous clinical pelvic inflammatory disease	13 (2)	12 (1)
Previous documented <i>Chlamydia trachomatis</i> infection	16 (2)	13 (1)
Mean (SD) No of pregnancies	2.0 (2)	1.2 (1.6)
Menopausal	119 (14)	144 (8)
First smear	0	220 (13)
Previous abnormal smears	828 (100)	68 (5)
Contraception:		
Combined oral contraceptive	368 (44)	913 (52)
Intrauterine device	90 (11)	176 (10)
Condoms	49 (6)	53 (3)
HIV positive	10 (1)	5 (0)
Centre:		
Hôpital Jean Verdier (public, Paris suburbs)	35 (4)	70 (4)
Hôpital Louis Mourier (public, Paris suburbs)	238 (29)	54 (3)
Laboratoire Cartier (private, Paris)	520 (63)	559 (32)
Centre d'Anatomie pathologique (private, Besançon)	35 (4)	1074 (61)
Results of colposcopy*:		
Normal	131 (16)	1405 (80)
Acetowhite epithelium	389 (47)	146 (8)
Iodine negative epithelium	665 (80)	322 (18)

\*Acetowhite and iodine negative epithelium are not mutually exclusive so the total exceeds 100%.

reading smears: all had extensive experience in conventional smears and cervical biopsies, but their experience with monolayer cytology was initially limited. However, this bias was neutralised by the optimised interpretations in which the best assessment was obtained.

### Implications

This study has implications for regulation of medical devices, clinical practice, and future research on screening for cervical cancer. Monolayer testing, which seems less reliable and less valid and is more expensive,<sup>3</sup> should not replace conventional smear tests for cervical cancer screening. Human papillomavirus testing as complementary to conventional smear testing should be further evaluated in clinical research.<sup>24</sup> Our results emphasise the need to improve the “hard evidence” in studies of new technologies for cervical screening by using adequate methodological standards.

Members of the French Society of Clinical Cytology Study Group: S Arkwright, A Biaggi, C Besançon-Roux, L Carbillon, I Cartier, B Cochand-Priollet (clinical coordinator), J Coste (methodological coordinator), J Darondel, P Dauvergne, P de Cremoux, A Dosda, E Foucher, I Gouget, D Grondard, B Karkouche, S Labbé, C Le Galès, I Le Guen, V Lepoutre, N Lestrat, H Magdelenat, A Malvy-Nickles, E Merea, V Molinié, A Odier, A Petitjean, S Peschard, P Piquet, L Pommaret, X Sastre-Garau, V Saha, N Seince, C Sigal-Hummel, M C Vacher-Lavenu, P Vielh, M Ziol.

Contributors: JC, M-CV-L, VM, BC-P, CLG, and PV initiated the study and developed the study protocol. JC and BCP coordinated the study. BCP, PdeC, IC, VM, M-CV-L, SL, and PV participated in coordinating the study, data collection, and edit-

**Table 2** Specimen adequacy and causes of inadequacy and limitation by method (clinical reading)\*

	Conventional cervical smear (n=2582)*	Monolayer (n=2580)†
Satisfactory for evaluation	2343 (91%)	2241 (87%)‡
Satisfactory for evaluation but limited by§:	236	328
Cytolysis	6	19
Obscuring blood	36	25
Obscuring inflammation	79	44
Endocervical component absent	75	235
Scant squamous epithelial cells	3	4
Air drying artefacts	32	0
Staining problem	5	1
Unsatisfactory for evaluation:	3	11
Cytolysis	0	0
Obscuring blood	0	5
Obscuring inflammation	0	1
Endocervical component absent	0	0
Scant squamous epithelial cells	1	5
Air drying artefacts	2	0
Staining problem	0	0

\*Data missing data for three.

†Data missing for five.

‡P<0.0001 for difference between groups (MacNemar  $\chi^2$  test).

§Causes of limited evaluation were not mutually exclusive.

ing. JC calculated the sample size and designed and performed the analysis. JC prepared the manuscript with input from all other authors, revised the manuscript for resubmission, and is the guarantor.

Funding: Direction Générale de la Santé and Programme Hospitalier de Recherche Clinique, French Ministry of Health

**Table 3** Interpretation of conventional cervical smear tests\* or monolayer testing versus† reference standard (colposcopy and biopsy) by population

Reference standard results	Women referred for colposcopy					Women attending for routine smears				
	Normal	ASCUS/AGUS	LSIL	HSIL	Invasive cancer	Normal	ASCUS/AGUS	LSIL	HSIL	Invasive cancer
<b>Conventional cervical smear (optimised interpretation)</b>										
Normal colposcopy/negative biopsy	231	26	25	7	0	1510	77	54	11	0
CIN grade I	30	21	157	37	0	22	11	22	6	0
CIN grades II-III	14	8	19	230	7	4	5	6	20	0
Invasive cancer	0	2	0	6	7	1	1	0	3	1
<b>Monolayer (optimised interpretation)</b>										
Normal colposcopy/negative biopsy	220	24	35	9	0	1493	73	73	15	0
CIN grade I	37	23	158	25	0	22	9	18	12	0
CIN grades II-III	19	10	32	212	4	5	3	5	22	0
Invasive cancer	0	3	0	9	3	0	0	1	4	0
<b>Conventional cervical smear (clinical reading)</b>										
Normal colposcopy/negative biopsy	175	31	63	21	0	1550	59	37	6	1
CIN grade I	18	20	173	33	0	23	8	24	6	0
CIN grades II-III	9	6	35	218	10	5	4	9	18	0
Invasive cancer	1	0	0	3	11	1	1	0	2	1
<b>Monolayer (clinical reading)</b>										
Normal colposcopy/negative biopsy	166	34	59	27	0	1503	85	51	9	0
CIN grade I	29	29	157	30	0	26	9	15	11	0
CIN grades II-III	11	12	37	211	7	8	4	7	17	0
Invasive cancer	1	1	0	7	6	1	0	0	4	0

ASCUS/AGUS=atypical squamous cells/glandular cells of undetermined significance.

LSIL=low grade squamous intraepithelial lesions.

HSIL=high grade squamous intraepithelial lesions.

CIN=cervical intraepithelial neoplasia.

\*Data missing for three to five conventional slides.

†Data missing for 11 monolayer slides

(AOM 98010); Association pour la Recherche sur le Cancer (9099). The guarantor accepts full responsibility for the conduct of the study, had access to the data, and controlled the decision to publish.

Competing interests: None declared.

Ethical approval: The approval of the ethics committee (Hôpital Cochin, Paris) for the study was obtained in July 1998.

- Nanda K, McCrory DC, Myers ER, Bastian LA, Hasselblad V, Hickey JD, et al. Accuracy of the Papanicolaou test in screening for and follow-up of cervical cytologic abnormalities: a systematic review. *Ann Intern Med* 2000;132:810-9.
- Hartmann KE, Nanda K, Hall S, Myers E. Technologic advances for evaluation of cervical cytology: is newer better? *Obstet Gynecol Surv* 2001;56:765-74.

- Merea E, Le Galès C, Cochand-Priollet B, Cartier I, de Cremoux P, Vacher-Lavenu MC, et al. Cost of screening for cancerous and precancerous lesions of the cervix. *Diagn Cytopathol* 2002;27:251-7.
- Cochand-Priollet B, Le Galès C, de Cremoux P, Molinié V, Sastre-Garau X, Vacher-Lavenu MC, et al. Cost-effectiveness of monolayers and human papillomavirus testing compared to that of conventional Papanicolaou smears for cervical cancer screening: protocol of the study of the French Society of Clinical Cytology. *Diagn Cytopathol* 2001;24:412-20.
- Reid MC, Lachs MS, Feinstein AR. Use of methodological standards in diagnostic test research. Getting better but still not good. *JAMA* 1995;274:645-51.
- Staff A, Wilbanks GD. An international terminology of colposcopy: report of the Nomenclature Committee of the International Federation of Cervical Pathology and Colposcopy. *Obstet Gynecol* 1991;77:313-4.
- Egglin TK, Feinstein AR. Context bias. A problem in diagnostic radiology. *JAMA* 1996;276:1752-5.

**Table 4** Interpretation of conventional cervical smear testing\* versus monolayer tests† by population

Result with conventional cervical smear test	Result with monolayer test									
	Women referred for colposcopy					Women attending for routine smears				
	Normal	ASCUS/AGUS	LSIL	HSIL	Invasive cancer	Normal	ASCUS/AGUS	LSIL	HSIL	Invasive cancer
<b>Optimised interpretation</b>										
Normal	229	19	17	7	0	1445	51	32	7	0
ASCUS/AGUS	19	17	16	3	0	48	21	15	9	0
LSIL	15	17	160	9	0	20	11	45	6	0
HSIL	12	7	32	227	2	4	2	5	29	0
Invasive cancer	0	0	0	9	5	0	0	0	1	0
<b>Clinical reading</b>										
Normal	147	17	26	9	0	1480	66	23	3	0
ASCUS/AGUS	24	14	18	1	0	39	20	9	4	0
LSIL	30	34	181	26	0	15	11	36	8	0
HSIL	6	9	28	230	2	1	1	5	25	0
Invasive cancer	0	1	0	9	11	1	0	0	1	0

ASCUS/AGUS=atypical squamous cells/glandular cells of undetermined significance. LSIL=low grade squamous intraepithelial lesions. HSIL=high grade squamous intraepithelial lesions.

\*Data missing for three to five conventional slides.

†Data missing for 11 monolayer slides.

**Table 5** Sensitivity, specificity, and likelihood ratios of conventional cervical smear testing, monolayer, and human papillomavirus (HPV) DNA testing for detection of cervical intraepithelial neoplasia grade I and above

Abnormality threshold	Colposcopy			Screening		
	Sensitivity(95% CI)	Specificity(95% CI)	Likelihood ratio+/likelihood ratio-	Sensitivity(95% CI)	Specificity(95% CI)	Likelihood ratio+/likelihood ratio-
<b>≥ASCUS/AGUS*</b>						
Cervical smear (clinical reading)	95 (93 to 97)	60 (55 to 66)	2.39/0.09	72 (63 to 80)	94 (93 to 95)	11.49/0.30
Monolayer (clinical reading)	92 (90 to 95)	58 (52 to 64)	2.19/0.13	66 (56 to 75)	91 (90 to 93)	7.47/0.38
Cervical smear (optimised interpretation)	92 (90 to 94)	80 (75 to 85)	4.58/0.10	74 (66 to 83)	91 (90 to 93)	8.64/0.28
Monolayer (optimised interpretation)	90 (87 to 92)	76 (71 to 81)	3.79/0.14	73 (65 to 82)	90 (89 to 92)	7.53/0.30
<b>≥Low grade squamous intraepithelial lesions†</b>						
Cervical smear (clinical reading)	90 (87 to 92)	71 (66 to 76)	3.11/0.14	59 (49 to 68)	97 (97 to 98)	22.10/0.42
Monolayer (clinical reading)	85 (82 to 88)	70 (65 to 75)	2.81/0.22	53 (43 to 63)	96 (95 to 97)	14.54/0.49
Cervical smear (optimised interpretation)	86 (83 to 89)	89 (85 to 93)	7.77/0.16	57 (48 to 67)	96 (95 to 97)	14.59/0.44
Monolayer (optimised interpretation)	83 (80 to 86)	85 (81 to 89)	5.42/0.20	61 (52 to 71)	95 (94 to 96)	11.54/0.41
<b>RLU/cut-off value ratio &gt;1.0‡</b>						
HPV DNA (high risk types)	79 (74 to 83)	77 (71 to 83)	3.46/0.28	64 (53 to 76)	86 (85 to 88)	4.73/0.41
HPV DNA (high and low risk types)	82 (77 to 86)	74 (67 to 80)	3.13/0.25	69 (58 to 79)	83 (81 to 85)	4.09/0.38
<b>≥Low grade squamous intraepithelial lesions or RLU/cut-off value ratio &gt;1.0 if ASCUS/AGUS¶</b>						
Monolayer (clinical reading) or HPV DNA (high risk types) if ASCUS/AGUS	87 (83 to 91)	66 (58 to 73)	2.41/0.20	59 (47 to 70)	96 (95 to 97)	13.50/0.43
Monolayer (optimised interpretation) or HPV DNA (high risk types) if ASCUS/AGUS	85 (80 to 89)	82 (76 to 88)	4.75/0.19	67 (56 to 78)	94 (92 to 95)	10.86/0.35

\*Clinical reading: conventional smear superior to monolayer for both sensitivity (p<0.05) and specificity (P<0.001). Optimised interpretation: conventional smear superior to monolayer for specificity (P<0.05) but not for sensitivity (P=0.07).

†Clinical reading: conventional smear superior to monolayer for sensitivity (P<0.001) but not for specificity (P=0.07). Optimised interpretation: conventional smear superior to monolayer for specificity (P<0.001) but not for sensitivity (P=0.08).

‡HPV results expressed as relative light units (RLUs), which represent ratio of light emission of sample to average of three positive controls provided by manufacturer, containing 1 pg/ml HPV 16 DAN. RLU ≥1&comma; corresponding to ≥5000 HPV-DNA copies per test well, was considered to be positive.

§Sensitivity and specificity of HPV DNA testing (high risk types) lower (P<0.0001) than both cytological techniques.

¶Clinical reading: paired monolayers/HPV testing superior but not significantly to standard conventional smear for sensitivity (P=0.88) and inferior for specificity (P<0.001). Optimised interpretation: paired monolayers/HPV testing superior but not significantly to standard conventional smear for sensitivity (P=0.06) and inferior for specificity (P<0.001).



**Table 6** Sensitivity, specificity, and likelihood ratios of conventional cervical smear testing, monolayer, and human papillomavirus (HPV) DNA testing for detection of cervical intraepithelial neoplasia grade II and above

Abnormality threshold	Colposcopy			Screening		
	Sensitivity(95% CI)	Specificity(95% CI)	Likelihood ratio+/ likelihood ratio-	Sensitivity(95% CI)	Specificity (95% CI)	Likelihood ratio+/ likelihood ratio-
<b>≥High grade squamous intraepithelial lesions *</b>						
Cervical smear (clinical reading)	83 (78 to 87)	90 (87 to 92)	8.17/0.19	51 (36 to 67)	99(99 to 100)	67.53/0.49
Monolayer (clinical reading)	79 (74 to 84)	89 (87 to 92)	7.34/0.24	51 (36 to 67)	99 (98 to 99)	43.25/0.49
Cervical smear (optimised interpretation)	85 (81 to 89)	92 (89 to 94)	10.36/0.16	60 (45 to 75)	99(99 to 99)	60.46/0.40
Monolayer (optimised interpretation)	78 (73 to 83)	94 (92 to 96)	12.19/0.23	65 (50 to 80)	98 (98 to 99)	41.29/0.36
<b>RLU/cut-off value ratio &gt;1.0†</b>						
HPV DNA (high risk types)	80 (74 to 86)	54 (49 to 60)	1.75/0.37	96 (88 to 100)	85 (83 to 87)	6.52/0.05
HPV DNA (high and low risk types)	81 (75 to 87)	50 (44 to 55)	1.60/0.39	96 (88 to 100)	82 (80 to 84)	5.32/0.05
<b>≥High grade squamous intraepithelial lesions or RLU/cut-off value ratio &gt;1.0 if ASCUS/AGUS‡</b>						
Monolayer (clinical reading) or HPV DNA (high risk types) if ASCUS/AGUS	82 (76 to 88)	86 (82 to 90)	5.94/0.21	64 (45 to 83)	98 (97 to 99)	28.47/0.37
‡Monolayer (optimised interpretation) or HPV DNA (high risk types) if ASCUS/AGUS	80 (74 to 86)	93 (90 to 96)	11.60/0.21	76 (59 to 93)	97(97 to 98)	29.71/0.25

\*Clinical reading: conventional smear is superior to monolayer but not significantly for both sensitivity (P=0.12) and specificity (P=0.16). Optimised interpretation: conventional smear is superior to monolayer for sensitivity (P<0.001) but not for specificity (P=1.00).

†Sensitivity of HPV DNA testing (high risk types) is higher but not significantly than that of conventional smear (P=0.66) but specificity is lower (P<0.0001).

‡Clinical reading: paired monolayers/HPV testing is superior but not significantly to standard conventional smear for sensitivity (P=0.29) but inferior for specificity (P<0.001). Optimised interpretation: paired monolayers/HPV testing is superior but not significantly to standard conventional smear for sensitivity (P=0.73) but inferior for specificity (P<0.001).

### What is already known on this topic

New technologies have been developed to improve the detection of cervical cancer and its precursors and reduce the rate of false negative results from conventional cervical smear tests

In several countries liquid based monolayer cytology is replacing conventional smear tests, despite controversy about whether these more expensive tests perform better

### What this study adds

Conventional cervical smear testing is superior in terms of low and high grade lesions and in populations with a low or a high incidence of abnormalities

Monolayer testing is less reliable and should not replace conventional cervical smear testing

- The Bethesda system for reporting cervical/vaginal cytologic diagnoses: revised after the second National Cancer Institute Workshop, April, 29-30, 1991. *Acta Cytol* 1993;37:115-24.
- Poljak M, Brenic A, Seme K, Vince A, Marin IJ. Comparative evaluation of first- and second-generation digene hybrid capture assays for detection of human papillomaviruses associated with high or intermediate risk for cervical cancer. *J Clin Microbiol* 1999;37:796-7.
- Fleiss J. *Statistical methods for rates and proportions*. New York: Wiley, 1981.
- Hutchinson ML, Zahniser DJ, Sherman ME, Herrero R, Alfaro M, Bratti MC, et al. Utility of liquid-based cytology for cervical carcinoma screening: results of a population-based study conducted in a region of Costa Rica with a high incidence of cervical carcinoma. *Cancer* 1999;87:48-55.
- Sheets EE, Constantine NM, Dimisco S, Dean B, Cibas ES. Colposcopically directed biopsy provide a basis for comparing the accuracy of ThinPrep and Papanicolaou smears. *J Gynecol Tech* 1995;1:27-33.
- Bishop JW, Bigner SH, Colgan TJ, Husain M, Howell LP, McIntosh KM, et al. Multicenter masked evaluation of AutoCyte PREP thin layers with matched conventional smears. Including initial biopsy results. *Acta Cytol* 1998;42:189-97.
- Roberts JM, Gurley AM, Thurlow JK, Bowditch R, Laverty CRA. Evaluation of the Thin Prep pap test as an adjunct to the conventional Pap smear. *Med J Aust* 1997;167:466-9.
- Bolick DR, Hellman DJ. Laboratory implementation and efficacy assessment of the ThinPrep cervical cancer screening system. *Acta Cytol* 1998;42:209-13.

**Table 7** Interobserver reliability for conventional smear tests and monolayer tests\*

First interpretation	Second interpretation				
	Normal	ASCUS/AGUS	LSIL	HSIL	Invasive cancer
<b>Cervical smear</b>					
Normal	448	35	15	13	0
ASCUS/AGUS	13	25	14	4	0
LSIL	34	11	33	8	0
HSIL	6	4	4	86	0
Invasive cancer	0	0	0	0	8
<b>Monolayer cytology</b>					
Normal	442	33	14	8	0
ASCUS/AGUS	29	11	10	6	0
LSIL	48	10	28	10	0
HSIL	16	10	22	65	0
Invasive cancer	0	0	0	0	2

\*Data available (two blinded interpretations) for 761 conventional smear slides and 764 monolayer slides.

- Park IA, Lee SN, Chae SW, Park KH, Kim JW, Lee HP. Comparing the accuracy of ThinPrep Pap tests and conventional Papanicolaou smears on the basis of the histologic diagnosis: a clinical study of women with cervical abnormalities. *Acta Cytol* 2001;45:525-31.
- Manos MM, Kinney WK, Hurley LB, Sherman ME, Shieh-Ngai J, Kurman RJ, et al. Identifying women with cervical neoplasia: using human papillomavirus DNA testing for equivocal Papanicolaou results. *JAMA* 1999;281:1605-10.
- Ronnett BM, Manos MM, Ransley JE, Fetterman BJ, Kinney WK, Hurley LB, et al. Atypical glandular cells of undetermined significance (AGUS): cytopathologic features, histopathologic results, and human papillomavirus DNA detection. *Hum Pathol* 1999;30:816-25.
- Bergeron C, Jeannel D, Poveda J, Cassonnet P, Orth G. Human papillomavirus testing in women with mild cytologic atypia. *Obstet Gynecol* 2000;95:821-7.
- Schneider A, Hoyer H, Lotz B, Leistriza S, Kuhne-Heid R, Nindl I, et al. Screening for high-grade cervical intra-epithelial neoplasia and cancer by testing for high-risk HPV, routine cytology or colposcopy. *Int J Cancer* 2000;89:529-34.
- Ferenczy A, Robitaille J, Franco E, Arseneau J, Richart RM, Wright TC. Conventional cervical cytologic smears vs. Thin Prep smears. A paired comparison study of cervical cytology. *Acta Cytol* 1996;40:1136-42.
- Bergeron C, Bishop J, Lemarie A, Cas F, Ayivi J, Huynh B, et al. Accuracy of thin-layer cytology in patients undergoing cervical cone biopsy. *Acta Cytol* 2001;45:519-24.
- Bernstein SJ, Sanchez-Ramos L, Ndubisi B. Liquid-based cervical cytologic smear study and conventional Papanicolaou smears: a metaanalysis of prospective studies comparing cytologic diagnosis and sample adequacy. *Am J Obstet Gynecol* 2001;185:308-17.
- Kaufman RH. Is there a role for human papillomavirus testing in clinical practice? *Obstet Gynecol* 2001;98:724-5. (Accepted 29 January 2003)